

GeneScan™ 500 LIZ® Size Standard

Catalog Number 4322682

Pub. no. 4363115 Rev. C

Contents†	Size	Storage conditions
LIZ® dye-labeled, single-stranded DNA fragments	2 X 200 µL (800 reactions‡)	Store up to one year at 2–8°C. Do not freeze.

† With the retirement of the ABI PRISM® 377 DNA Sequencer this kit no longer contains the gel loading buffer. If you need to continue to purchase the loading buffer, it is available as a separate part (Cat. no. 402055).

‡ 800 reactions when using the recommended loading amount of 0.5 µL.

WARNING! Read the Safety Data Sheets (SDSs) and follow the handling instructions. Wear appropriate protective eyewear, clothing, and gloves. Safety Data Sheets (SDSs) are available from www.lifetechnologies.com/support.

Product description

GeneScan™ 500 LIZ® Size Standard is an internal lane size standard developed for use with Life Technologies fluorescence-based DNA electrophoresis systems. The use of an internal lane size standard during electrophoresis enables automated data analysis and precise DNA fragment size comparisons between electrophoresis runs. GeneScan™ 500 LIZ® Size Standard is designed for sizing DNA fragments in the 35–500 bp range and provides 16 single-stranded labeled fragments of 35, 50, 75, 100, 139, 150, 160, 200, 250, 300, 340, 350, 400, 450, 490, and 500 bases. Each of the DNA fragments is labeled with the LIZ® fluorophore, which results in a single peak when run under denaturing conditions.

Instructions for use

Sample preparation

1. Before use, mix the contents of each tube thoroughly and centrifuge briefly to collect the liquid at the bottom of the tube. Typical loading cocktails are as follows:

Component	Electrophoresis system		
	310	3130/3100 Series	3730 Series
Sample	0.5 µL	0.5 µL	0.5 µL
GeneScan™ 500 LIZ® Size Standard	0.5 µL	0.25 µL	0.5 µL
Hi-Di™ Formamide (Cat. no. 4311320)†	9.0 µL	9.25 µL	9.0 µL

† Not included in this kit.

Note: We highly recommend using the ratios of DNA sample (PCR product) and size standard presented in the table as a starting point only. Optimize the ratios if necessary, based on your experimental results. Also, see guidelines in the following section.

2. Heat the loading cocktail for 3 minutes at 95°C.
3. Immediately chill on ice for a few minutes and load samples.

Guidelines



Optimize your analysis based on the following information about use of this standard on capillary electrophoresis instruments:

- The 250 bp peak is sensitive to small temperature variations. The 250 bp fragment should not be used when defining the size standard in GeneMapper® Software.
- The 340 bp peak is subject to large temperature variations.
- Fragment analysis primer peaks can often interfere with the detection of the 35 bp peak.

Note: Discard any unused reagent that has been diluted in Hi-Di™ Formamide.

Limited product warranty

Life Technologies Corporation and/or its affiliate(s) warrant their products as set forth in the Life Technologies' General Terms and Conditions of Sale found on Life Technologies' website at www.lifetechnologies.com/termsandconditions. If you have any questions, please contact Life Technologies at www.lifetechnologies.com/support.

For Research Use Only. Not for use in diagnostic procedures.

The information in this guide is subject to change without notice.

DISCLAIMER: LIFE TECHNOLOGIES CORPORATION AND/OR ITS AFFILIATE(S) DISCLAIM ALL WARRANTIES WITH RESPECT TO THIS DOCUMENT, EXPRESSED OR IMPLIED, INCLUDING BUT NOT LIMITED TO THOSE OF MERCHANTABILITY, FITNESS FOR A PARTICULAR PURPOSE, OR NON-INFRINGEMENT. TO THE EXTENT ALLOWED BY LAW, IN NO EVENT SHALL LIFE TECHNOLOGIES AND/OR ITS AFFILIATE(S) BE LIABLE, WHETHER IN CONTRACT, TORT, WARRANTY, OR UNDER ANY STATUTE OR ON ANY OTHER BASIS FOR SPECIAL, INCIDENTAL, INDIRECT, PUNITIVE, MULTIPLE OR CONSEQUENTIAL DAMAGES IN CONNECTION WITH OR ARISING FROM THIS DOCUMENT, INCLUDING BUT NOT LIMITED TO THE USE THEREOF.

NOTICE TO PURCHASER: DISCLAIMER OF LICENSE: This product is compatible with DNA sequencing or fragment analysis methods covered by patents owned or licensed by Applied Biosystems. No license under such patents and no right to resell or repackage this product in any form (whether alone or with other materials) or to provide services using this product for a fee is conveyed expressly, by implication, or by estoppel. A license to use the DNA sequencing or fragment analysis methods for certain research and development activities is available from Applied Biosystems. Further information on purchasing licenses may be obtained by contacting the Director of Licensing, Applied Biosystems, 850 Lincoln Centre Drive, Foster City, California 94404, USA.

Limited Use Label License No: 358: Research Use Only

The purchase of this product conveys to the purchaser the limited, non-transferable right to use the product only to perform internal research for the sole benefit of the purchaser. No right to resell this product or any of its components is conveyed expressly, by implication, or by estoppel. This product is for internal research purposes only and is not for use in commercial applications of any kind, including, without limitation, quality control and commercial services such as reporting the results of purchaser's activities for a fee or other form of consideration. For information on obtaining additional rights, please contact outlicensing@lifetech.com.

© 2012 Life Technologies Corporation. All rights reserved. The trademarks mentioned herein are the property of Life Technologies Corporation and/or its affiliates or their respective owners.

For support visit lifetechnologies.com/support or email techsupport@lifetech.com

lifetechnologies.com

13 December 2012

